



Certificate of Analysis

Verified by:
Jyothi Vishumangalam

J. Vishumangalam

Quality Release date:
May 10, 2018

Chai Biotechnologies Inc.
Santa Clara, CA 95050
Phone: +1 (650) 779-5577


Product Name: **Chai Green 20X**
Catalog Num: R01200
Lot Num: 7093515
Analysis Date: May 10, 2018
Expiration Date: May 10, 2020
Storage: Store at -20 °C

Test

Functional Testing*	Specification	Results
Lambda assay efficiency	90-110%	PASS
Lambda assay R ²	>0.990	PASS
Absorbance at pH 8		
O.D. 488 nM	0.1 < Value < 0.18	PASS

* Functional testing—Lambda assay efficiency and R² was determined using varying starting amounts of lambda phage genomic DNA as template, primers, Chai PCR Master Mix (2X) and Chai Green 20X. A 200 bp amplicon was amplified using cycling conditions of 3 min @ 95 °C, 40 x (15 s @ 95 °C, 60 s @ 60 °C).

Standard Curve



Lambda Assay using Chai Green

$$y = -3.304x + 13.039$$

$$R^2 = 0.997$$